



Allweighs Weighing Services
Westend Grange
West End Road
Epworth
Doncaster
DN9 1LA
Tel: 01427 875 867
Web: www.allweighs.co.uk



Certificate No: KCS:

202401066

Location:

Out weighbridge

Manufacturer:

Allweighs / Dini Argeo 3590 ET

Serial/Plant No:

25145165

Capacity:

50000

kg

Increment:

20

kg

Frequency:

12

months

Test Date:

04 October 2024

Next due:

October 2025

Calibration Certificate of Accuracy

Business Name:

Go4Greener

Address:

201 Slack Lane
Derby
Derbyshire
DE22 3EE

Contact Name:

Tim Richardson

Off-centre Loading:

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	6700	0	5	6700	0
2	6700	0	6	6700	0
3	6700	0			
4	6700	0			

☐ Four loadcell

☒ Six loadcell

☐ Single loadcell

☐ Eight loadcell

As Found

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
20700	20700	0
30700	30700	0
43960	43960	0

Repeatability

43960	43960	0
-------	-------	---

Notes:

FOUND TO BE CORRECT

Engineer
Name:

KS

Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
20700	20700	0
30700	30700	0
43960	43960	0

Repeatability

43960	43960	0
-------	-------	---

Traceable Weights Used:

W24017

Trading
Standards
Maximum
Tolerance:

60

+/-

60

kg

Calibration method and equipment used
traceable to national standards.

QMF17 Issue 5 14/07/2020